

**SPECIAL 510(k): Device Modification**  
**ODE Review Memorandum (Decision Making Document is Attached)**

**To:** THE FILE

**RE:** DOCUMENT NUMBER K042206

---

This 510(k) submission contains information/data on modifications made to the SUBMITTER'S own Class II, Class III or Class I devices requiring 510(k). The following items are present and acceptable (delete/add items as necessary):

1. The name and 510(k) number of the SUBMITTER'S previously cleared device. (For a preamendments device, a statement to this effect has been provided.) – Medtronic HMS Purple/Black Control - #K922031
2. Submitter's statement that the **INDICATION/INTENDED USE** of the modified device as described in its labeling **HAS NOT CHANGED** along with the proposed labeling which includes instructions for use, package labeling, and, if available, advertisements or promotional materials (labeling changes are permitted as long as they do not affect the intended use). This information is provided in pp. 2-3 of the submission.
3. A description of the device **MODIFICATION(S)**, including clearly labeled diagrams, engineering drawings, photographs, user's and/or service manuals in sufficient detail to demonstrate that the **FUNDAMENTAL SCIENTIFIC TECHNOLOGY** of the modified device **has not changed**.  
**This change was for** replacing USP bovine heparin with USP porcine heparin, because it is no longer commercially available for manufacture of clinical use.
4. **Comparison Information** (similarities and differences) to applicant's legally marketed predicate device including, labeling, intended use, physical characteristics, and storage/stability. The information is supplied on p. 3 in Table 1.
5. A **Design Control Activities Summary** which includes:
  - a) Identification of Risk Analysis method(s) used to assess the impact of the modification on the device and its components, and the results of the analysis. This information is found in Appendix C. The company submitted Failure Mode and Effects Analysis (FMEA), including a Hazards Checklist, using these documents:
    - PC1.48 – Product/Process Risk Management Procedure
    - ISO 14971 – Medical Devices; Application of Risk Management to Medical Devices
    - 21 CFR Part 820 - Quality Systems Regulations
    - 98/79EC – In vitro Diagnostic Device Directive
    - AP5514 – Product Structure, Purple/Black HMS Control
  - b) Based on the Risk Analysis, an identification of the verification and/or validation activities required, including methods or tests used and acceptance criteria to be applied. They also supplied test data for clotting times of (3) control lots in Channels 5/6 of cartridges (N=50). All (3) lots met the acceptance criteria of < 249 seconds. This information is found in Appendix D.
  - c) A declaration of conformity with design controls. The declaration of conformity should include:
    - i) A statement signed by the individual responsible, that, as required by the risk analysis, all verification and validation activities were performed by the designated individual(s) and the results demonstrated that the predetermined acceptance criteria were met, and
    - ii) A statement signed by the individual responsible, that the manufacturing facility is in conformance with design control procedure requirements as specified in 21 CFR 820.30 and the records are available for review. The document is found in Appendix E.

6. A Truthful and Accurate Statement, a 510(k) Summary or Statement and the Indications for Use Enclosure (and Class III Summary for Class III devices). These documents are found in Appendices H, F and E, respectively.

The labeling for this modified subject device has been reviewed to verify that the indication/intended use for the device is unaffected by the modification. In addition, the submitter's description of the particular modification(s) and the comparative information between the modified and unmodified devices demonstrate that the fundamental scientific technology has not changed. The submitter has provided the design control information as specified in The New 510(k) Paradigm and on this basis, I recommend the device be determined substantially equivalent to the previously cleared (or their preamendment) device.

	_____	_____
	(Reviewer's Signature)	(Date)
Comments		
_____		
_____		
_____		

revised:8/1/03