

**510(k) SUBSTANTIAL EQUIVALENCE DETERMINATION
DECISION SUMMARY
DEVICE TEMPLATE**

A. 510(k) Number: K031577

B. Analyte: Protein 2 Calibration verifiers A – E for Beckman Coulter Array Protein Systems containing the following analytes; Alpha-1-Acid Glycoprotein, Alpha-1-Antitrypsin, Alpha-2-Macroglobulin, Antithrombin III, Beta-2-Macroglobulin and Ceruloplasmin.

C. Type of Test: n/a

D. Applicant: CLINIQA

E. Proprietary and Established Names: Quality Control Material

F. Regulatory Information:

1. Regulation section: 21 CFR 862.1660
2. Classification: Class 1
3. Product Code: JJY
4. Panel: 75

G. Intended Use:

1. Indication(s) for use: The LiniCAL Protein 2 Calibration Verifiers Levels A – E for Beckman Coulter Array Protein Systems is intended for use as an assayed quality control material for analysis.
2. Special condition for use statement(s): none
3. Special instrument Requirements: none

H. Device Description: The Protein 2 Calibration Verifiers are used in the clinical laboratory to verify calibration and/or assess linearity of the Beckman Coulter Array Protein Systems. Five assayed levels of Alpha-1-Acid Glycoprotein, Alpha-1-Antitrypsin, Alpha-2-Macroglobulin, Antithrombin III, Beta-2-Macroglobulin and Ceruloplasmin are provided to allow monitoring of the reportable range.

I. Substantial Equivalence Information:

1. Predicate device name(s): LiniCAL Protein 2 Calibration Verifiers Levels A-E for Beckman Immage

2. Predicate K number(s): K023250
3. Comparison with predicate: Both devices are serum based products containing Alpha-1-Acid Glycoprotein, Alpha-1-Antitrypsin, Alpha-2-Macroglobulin, Antithrombin III, Beta-2-Macroglobulin and Ceruloplasmin and are manufactured using the same processes. The difference between the two products is the constituent concentration has been optimized for this test system.

J. Standard/Guidance Document Referenced (if applicable): N/A

K. Test Principle: N/A

L. Performance Characteristics (if/when applicable):

1. Analytical performance:
 - a. *Precision/Reproducibility*: N/A
 - b. *Linearity/assay reportable range*: N/a
 - c. *Traceability (controls, calibrators, or method)*: None provided
 - d. *Detection limit (functional sensitivity)*: N/A
 - e. *Analytical specificity*: N/A
 - f. *Assay cut-off*: N/A
2. Comparison studies:
 - a. *Method comparison with predicate device*: N/A
 - b. *Matrix comparison*: N/A
3. Clinical studies:
 - a. *Clinical sensitivity*: N/A
 - b. *Clinical specificity*: N/A
4. Clinical cut-off: N/A
5. Expected values/Reference range: N/A

M. Conclusion: Based upon the information provided, I recommend that the CliniqaliniCAL Protein 2 Calibration verifiers Levels A-E for Beckman Coulter Array Protein Systems be found substantially equivalent with predicate devices according to 21 CFR 862.1660.